

Barometric Pressure Sensor AB 60

- Piezoelectric barometric pressure sensor
- Low power consumption
- Operating pressure:
 AB 60 (S31105) 800 ... 1100 hPa (mbar)

Measurement principle and mounting

The piezoelectric pressure sensor's signal is electronically amplified to provide an output signal of 0...5 VDC. The sensor is mounted in a stainless steel housing, protection class IP64 when the connector is plugged in.

If required, the sensor is ready for connection in an Ammonit steel cabinet.

When mounted outside the central steel cabinet we recommend protective housing with pressure compensation.

In measurement operation the sensor needs an extern supply of at least 9 VDC.

Specifications


Characteristics	AB 60				
Order-No.	S31105				
Operating pressure	800 1100 hPa (mbar) (Altitude: ≤ 1400 m)				
Slope	60 hPa/V				
Offset	800 hPa				
Temperature operation range	-40 85 °C				
Humidity range	0 98 %RH				
Accuracy					
Total accuracy (-10 60 °C)	±3 hPa				
Repeatability	±0,6 hPa				
Long term stability	±0,3 hPa/year				
Electrical data					
Output voltage	0 5 VDC				
Supply voltage	9 32 V				
Current consumption	5 mA				
General					
Dimensions	Length 72 mm, diameter 22 mm				
Weight	80 g				
Housing	Stainless steel				
Connection	4-pole plug (M12)				
Protection class	IP 64 - when connector is plugged in				
Vibration (5 500 Hz)	2 gRMS				
Mechanical shock	50 g				
Atmosphere	non-ionic, non-corrosive				

^{*} FSO (Full Scale Output) describes the difference of the upper and the lower limit of the pressure range.



S31105

Dimensional drawing



mass: appr. 80g dimensions in mm


Sensor connection to Ammonit Meteo-40 data logger

Sensor	Plug Pin No.	Ammonit Cable Wire Colour	Meteo-40 Analog Voltage	Supply Sensor
Air Pressure Output Voltage	2	white	Ах А	
Ground	4	blue	Ахв	
Supply	1	red		9 32 VDC
Ground	4	black		Main Ground

Cable type: LiYCY 4 x 0.25 mm²

Connect the shield logger-sided to Ground (GND)

Sensor connection diagram to Ammonit Meteo-40 data logger



page 2/2

Last Modification: 06 February 2023